



Item number 4430



Classic

Examination gloves, ABENA Classic, XL, clear, vinyl, powder free

- ✓ Smooth surface
- ✓ No latex proteins or Accelerators



Product Description

Examination glove with smooth surface and contains no latex proteins or accelerators. Suitable for handling non-fatty food products and brief cleaning tasks without strong chemicals.



LOW GRIP



EN374-5:2016



374-1:2016/
type B



AQL 1.5 EN EN 420 455

Specifications

Base name	Examination gloves
Brand	ABENA
Sub-brand	Classic
Size	XL
Properties	Latex free, low grip, rolled cuff, smooth
Features	Powder free
Single or multiple use	Single use
Material	Vinyl
Technical ingredients	PVC, DINP, Viscosity reducer JNJ, PU-coating, Ca-Zn Stabilizer
Length/depth	240 mm
Width	115 mm
Thickness	Min. 0,07 mm
Certifications	CE, Food contact materials, CAT III, MD
CE Category (Personal Protective Equipment)	CAT III
CE Class (Medical Devices)	Class I
Product or test standards	EN 455, AQL 1.5, EN 420, EN ISO 374-1:2016 Type B KPT, EN 374-5:2016 VIRUS
Directives, regulations and acts	(EU) No 10/2011, (EC) No 2023/2006, (EU) 2016/425, BEK nr 681 af 25/05/2020, (EC) No 1935/2004, MDR (EU) 2017/745
Safety Instructions and Warnings	Avoid contact with fatty food products as the plasticizers in the glove can migrate to food during prolonged contact. Contain PVC and plasticizers.
Because of the products composition it has an expiry date from the production date on:	5 years
Storage Instructions	Store dry, clean and at room temperature.
Product Disposal Instructions	Can be disposed of with normal household waste sorted according to local regulations.
Packaging Disposal Instructions	Can be disposed of with normal household waste sorted according to local regulations.

Instructions for use/application

Examine the gloves for flaws and defects before use

Packaging data

Unit	Contains	Length	Width	Height	EAN
cII	1000 pcs	36 cm	26 cm	28 cm	5703538137235
pck	100 pcs	24.5 cm	13 cm	6.5 cm	5703538137211
pcs	1 pcs				



Regulation (EU) 2017/745 is a regulation of the European Union on the clinical investigation and sale of medical devices for human use. It repeals Directive 93/42/EEC, which concerns medical devices, and Directive 90/385/EEC, which concerns active implantable medical devices, on 26 May 2021.



The CE mark guarantees that a product is safe to use and complies with all safety precautions. CE stands for Conformité Européenne (European Conformity) and is mainly found on electronic equipment, safety equipment, construction products, and medical devices.



The glass-fork-symbol guarantees that products live up to the demands stated in the European legislation for food contact materials. The symbol is not mandatory but advised on products used for food contact.



Third-party type approval is required for all Category III personal protective equipment (PPE) products. CAT III products will be accompanied by a notified body number for the 3rd party verification of the product's safety.



The product does not contain latex.



The glove has a smooth surface.



The glove has a low grip.



The glove is 240 mm long.

EN374-5:2016



The standard specifies the requirements for protective gloves against hazardous chemicals and microorganisms. This part of the standard describes the glove's resistance to bacteria, fungi and viruses.



The standard specifies the requirements for protective gloves against chemicals and microorganisms. Type B has been tested for a permeation time of at least 30 minutes with the chemicals sodium hydroxide 40%, hydrogen peroxide 30% and formaldehyde 37%.



The standard specifies the requirements for protective gloves in relation to general requirements and test methods.



EN 455 consists of four standards that gloves must be tested against in order to be a medical glove. This standard specifies the requirements for disposable medical gloves.

